



UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

Ph: 0522 2838102 e-mail: drugs@upmsc.in

Result of Technical and Sample Evaluation of Tender no. UPMSC/Drugs-085/289 (Mask (N-95) without valve) Dated: 02.03.2021

S. No	Name of Firm	Technically Qualified/ Not qualified	Reason for technical disqualification	Status of Sample	Committee observation for samples	Reason for Sample disqualification	Final result (based on Technical and Sample Evaluation)
1.	Amit Gramodyog Sansthan	Not eligible	-Firm has not submitted copy of original certificate issued by DRDE/BIS as per corrigendum-1	Received	Recommended	-	Not eligible
2.	D.K. Power Engineers Private Limited	Eligible	-	Received	Recommended	-	Eligible
3.	Irvine Technologies Pvt. Ltd.	Not eligible	-Firm has not submitted copy of original certificate issued by NIOSH-42C FR84/ EN149-2001 for N-95 mask as per corrigendum-1. -Non – Conviction certificate (self declaration) is not submitted. -Product permission copy is not legible. -Format – XVII (committed quantity) is not submitted.	Received	Not recommended	Due to presence of ear loop in place of head band.	Not eligible
4.	Pragati Meditech	Not eligible	Turnover for last 3 FY certificate should be submitted as per tender condition. Firm has submitted unaudited provisional turnover of FY 2020-21 Rs-3.07 crore.	Received	Recommended	-	Not eligible
5.	Sanmati Packaging	Eligible	-	Received	Recommended	-	Eligible
6.	Med Freshe Private Limited	Eligible	-	Received	Recommended	-	Eligible
7.	Nandani Creation Limited	Eligible	-	Received	Recommended	-	Eligible
8.	Neelkanth Packaging Solution	Eligible	-	Received	Recommended	-	Eligible
9.	Newtech Medical Devices	Not eligible	-Firm has not submitted copy of original certificate issued by NIOSH-42C FR84/ EN149-2001/ DRDE/BIS for N-95 mask without valve as per corrigendum-1. -Product permission is not as per tender.	Received	Recommended	-	Not eligible
10.	Paramount Surgimed Ltd	Not eligible	-Firm has not submitted copy of original certificate issued by DRDE/BIS N-95 mask without valve as per corrigendum-1.	Received	Recommended	-	Not eligible
11.	Rashmi Rare Earth Limited	Not eligible	Firm has not submitted turnover certificate of FY 2019-20.	Received	Recommended	-	Not eligible

S. No	Name of Firm	Technically Qualified/ Not qualified	Reason for technical disqualification	Status of Sample	Committee observation for samples	Reason for Sample disqualification	Final result (based on Technical and Sample Evaluation)
12.	Sanson Promo Innovations Llp	Eligible	-	Received	Recommended	-	Eligible
13.	Shakti Engineering Industries	Not eligible	-Firm has not submitted copy of certificate issued by DRDE/BIS for N-95 mask as per corrigendum-1.	Received	Recommended	-	Not eligible
14.	Shasti Spuntec India Private Limited	Not eligible	-Firm has not submitted copy of original certificate issued by DRDE/BIS as per corrigendum-1. -Non – Conviction certificate (self declaration) is not submitted. -60 days production capacity certificate is not submitted. -Turnover certificate of FY 2019-20 is not submitted	Received	Recommended	-	Not eligible
15.	Suraksha Kavach Limited	Not eligible	-Format – IV is not submitted. -Firm has not submitted copy of original certificate issued by NIOSH-42C FR84/ EN149-2001/ DRDE/BIS for N-95 mask as per corrigendum-1. -Non – Conviction certificate (self declaration) is not submitted. -Format – V is not submitted. -60 days production capacity certificate is not submitted. -Format – IX is not submitted. -Turnover certificate of FY 2017-18, 2018-19 & 2019-20 is not submitted. -UTR no of Tender fee and EMD are not identified.	Received	Recommended	-	Not eligible
16.	Unishield Corporation	Not eligible	Improper document has been submitted by firm.	Received	Not recommended	Due to presence of ear loop in place of head band.	Not eligible
17.	Unity Healthcare	Not eligible	Improper document has been submitted by firm.	Received	Recommended	-	Not eligible

Managing Director

UPMSCL