



UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

Ph: 0522 2838102 e-mail: drugs@upmsc.in

Provisional List of bidders/items of tender no. UPMSCL/Drugs-105/432 Date: 04 May, 2021

In reference to below observations, if any firm have a clarification/representation regarding any firm/items, then it must be communicated to UPMSCL through email i.e. drugs@upmsc.in before **15:00 PM, 05 May, 2021**. No communication thereafter in this regard shall be entertained by UPMSCL.

This is to inform that those bidders who have not submitted physical samples against above said tender may submit samples to UPMSCL Head office, SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010, **latest by 15:00 Hrs, 05 May, 2021**.

S N	Name of Firm	Observation Related To Firm	Provisional Status of the firm	Status of Sample
1	Bulk MRO Industrial Supply Private Limited	<ul style="list-style-type: none">➤ Firm has not submitted valid ISO/Quality Management System Certificate/BIS/Any other quality certificates issued by competent authority as per tender.➤ Firm has not submitted quality compliant with the below standards, or equivalent:<ul style="list-style-type: none">- EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374- ANSI/SEA 105-2011- ASTM D6319-10➤ Firm has not submitted test report of sample batch from Government laboratory/NABL➤ Firm has not submitted Non- Conviction certificate (self declaration)➤ Firm has not submitted Format – V➤ Firm has not submitted copy of the valid license to manufacture/import the item(s) quoted as per specifications mentioned in the tender from the Competent Authority or DIC or FDA as per tender.➤ Firm has not submitted list of Govt. Organizations to whom bidder is an existing Supplier.➤ Firm has not submitted Format-VII➤ Firm has not submitted EMD (MSME other State)	Not – Eligible	Not Received
2	UNITY HEALTHCARE	<ul style="list-style-type: none">➤ Firm has not submitted quality compliant with the below standards, or equivalent:<ul style="list-style-type: none">-EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374- ANSI/SEA 105-2011- ASTM D6319-10➤ Firm has not submitted copy of the valid license to manufacture/import the item(s) quoted as per specifications mentioned in the tender from the Competent Authority or DIC or FDA as per tender.	Not – Eligible	Received

Managing Director

UPMSCL