



UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

Ph: 0522 2838102 e-mail: drugs@upmsc.in

Provisional List of bidders/items of tender no. UPMSCL/Drugs-112/459 Date: 22.06.2021

In reference to below observations, if any firm have a clarification/representation regarding any firm/items, then it must be communicated to UPMSCL through e-mail i.e. drugs@upmsc.in before 17:00 PM, 23.06.2021. No communication thereafter in this regard shall be entertained by UPMSCL.

S. No	Name of Firm	Observation Related To Firm/Items	Provisional Status of the firm	Status of Sample
1	M/s Hi Care Gloves Pvt Ltd	<p>-MSME and DIC certificate are not legible.</p> <p>-Firm has not submitted copy of Quality compliant with the below standards, or equivalent:</p> <ul style="list-style-type: none">- EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374- ANSI/SEA 105-2011- ASTM D6319-10 <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS13422) for Hi care Gloves Pvt Ltd & submitted IS13422 is issued for Neerakkal Gloves Pvt Ltd</p> <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354).</p> <p>-Firm has not submitted 60 days production capacity for nitrile gloves</p> <p>-Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves and rubber examination gloves.</p> <p>-Firm has not submitted average annual turnover certificate with audited /provisional balance sheet for the FY 2020-21.</p>	Not-Eligible	Submitted
2	M/s Jajoo Surgicals Pvt. Ltd.	<p>-Firm has not submitted copy of authorization letter from Importer.</p> <p>-Firm has not submitted copy of Quality compliant with the below standards, or equivalent:</p> <ul style="list-style-type: none">- EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374- ANSI/SEA 105-2011- ASTM D6319-10 <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354).</p> <p>-Firm has not submitted copy of test report from Government laboratory/NABL for nitrile gloves and rubber examination gloves.</p>	Not-Eligible	Submitted
3	M/s M/s Sanmati Packaging	<p>-Firm has not submitted copy of Quality compliant with the below standards, or equivalent:</p> <ul style="list-style-type: none">- EU standard directive 93/42/EEC Class I, EN 455- EU standard directive 89/686/EEC Category III, EN 374- ANSI/SEA 105-2011- ASTM D6319-10 <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354).</p> <p>-Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS13422)</p>	Not-Eligible	Submitted

		-Firm has not submitted copy of authorization letter from Importer. -Firm has not submitted average annual turnover certificate with audited /provisional balance sheet for the FY 2020-21.		
4	M/s National Meditek	-Firm has not submitted average annual turnover certificate with audited /Provisional balance sheet for the FY 2020-21.	Not-Eligible	Submitted
5	M/s RFB Latex Limited	-Firm has not submitted copy of Quality compliant with the below standards, or equivalent: - EU standard directive 93/42/EEC Class I, EN 455 - EU standard directive 89/686/EEC Category III, EN 374 - ANSI/SEA 105-2011 - ASTM D6319-10 -Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS13422) -Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves and sterile surgical gloves pair.	Not-Eligible	Submitted
6	M/s Sky Lifecare Pvt Ltd	-As per tender specification, Outer gloves preferably reach mid-forearm (minimum 280 mm total length) while firm is quoted for 240 mm only. -Firm has not submitted copy of Quality compliant with the below standards, or equivalent: - EU standard directive 93/42/EEC Class I, EN 455 - EU standard directive 89/686/EEC Category III, EN 374 - ANSI/SEA 105-2011 - ASTM D6319-10 -Firm has not submitted copy of valid BIS certificate for the gloves (confirms to IS 15354). -Firm has not submitted copy of authorization letter from Importer. -Firm has not submitted copy of test report from Government laboratory/NABL for Nitrile gloves and rubber examination gloves. -Firm has not submitted average annual turnover certificate with audited /provisional balance sheet for the FY 2020-21.	Not-Eligible	Submitted
7	M/s Sogo Fashions Private Limited	-Firm has not submitted average annual turnover certificate with audited /Provisional balance sheet for the FY 2020-21.	Not-Eligible	Submitted
8	M/s Unity Healthcare	-Firm has not submitted average annual turnover certificate with audited /Provisional balance sheet for the FY 2020-21.	Not-Eligible	Submitted

Managing Director
UPMSCL