

UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010 Ph: 0522 2838102 e-mail: drugs@upmsc.in

Provisional List of bidders/items of tender no. UPMSCL/Drugs-075/253 Date: 09.10.2020

In reference to below observations, if any firm have a clarification/representation regarding any firm/items, then it must be communicated to UPMSCL through email i.e. drugs@upmsc.in before 17:00 PM, 12 October, 2020. No communication thereafter in this regard shall be entertained by UPMSCL.

This is to inform that those bidders who have not submitted physical samples of Triple Layer Mask, against above said tender may submit samples of Triple Layer mask to UPMSCL Head office, SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010, latest by 17:00 PM, 12 October, 2020.

S.No	Name of Firm	Observation Related To Firm	Provisional Status of the firm	Status of Sample
1	ASSOCIATES NONWOVENS	- Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted	Not – Eligible	Received
2	Deepali Designs and Exhibits Pvt Ltd	-	Eligible	Received
3	DYS IMPEX PRIVATE LIMITED	- No technical document submitted by bidder	Not – Eligible	Not – Received
4	jind surgical	- Test report from govt. laboratory/NABL bid for filtration efficiency (>99%) of the mask is not submitted - 60 days production capacity certificate not submitted	Not – Eligible	Received
5	K.S. Surgical Pvt Ltd	- Test report from govt. laboratory/NABL bid for filtration efficiency (>99%) of the mask is not submitted - Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted	Not – Eligible	Received
6	Manmohan Commercials Limited	- Format – XVII committed quantity not submitted - Test report from govt. laboratory/NABL bid for filtration efficiency (>99%) of the mask is not submitted - Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted	Not – Eligible	Received
7	MEDIBLUE HEALTHCARE PVT LTD	- Test report from govt. laboratory for filtration efficiency of the mask is not submitted as per tender requirement	Not – Eligible	Received

S.No	Name of Firm	Observation Related To Firm	Provisional Status of the firm	Status of Sample
		(i.e. >99%) - Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted		
8	Medicap Healthcare Limited	- Test report from govt. laboratory for filtration efficiency of the mask is not submitted as per tender requirement (i.e. >99%) - Manufacturing License is not submitted for Triple Layer Mask - Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted	Not – Eligible	Received
9	OM SURGICAL INDUSTRIES	- 60 days production capacity certificate not submitted	Not – Eligible	Received
10	P and O International Inc	- Format – XVII committed quantity not submitted - EMD and Tender fee not identified - Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted - 60 days production capacity certificate not submitted	Not – Eligible	Not – Received
11	paramount surgimed ltd	- Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted	Not – Eligible	Received
12	Rajshree Fabrics	- 60 days production capacity certificate not submitted	Not – Eligible	Received
13	Romsons Scientific And Surgical Industries Pvt. Ltd.	 Manufacturing License is not submitted for Triple Layer Mask Product permission is not as per tender Committed quantity is less than required quantity Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted 60 days production capacity certificate not submitted 	Not – Eligible	Received
14	Sanmati Packaging	 Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted 60 days production capacity certificate not submitted 	Not – Eligible	Received
15	Sanson Promo Innovations LLP	- Quality compliant certificate: ISO 13485/ISO 9001/EN 14683 or equivalent is not submitted - Test report from govt. laboratory/NABL bid for filtration efficiency (>99%) of the mask is not	Not – Eligible	Received

S.No	Name of Firm	Observation Related To Firm	Provisional Status of the firm	Status of Sample
		submitted (SITRA not in the name of firm) - 60 days production capacity certificate not submitted		
16	Shree Deepak Export	 Quality compliant certificate: ISO 13485/ISO 9001/EN 14683 or equivalent is not submitted Test report from govt. laboratory/NABL bid for filtration efficiency (>99%) of the mask not on the name of bidder Format – IV not submitted Committed quantity is less than required quantity Annual turnover certificate not certified by CA 60 days production capacity certificate not submitted 	Not – Eligible	Not – Received
17	Vikrant Life Sciences Pvt. Ltd.	- Annual turnover statement for FY - 2019 – 20 audited/provisional not submitted	Not – Eligible	Received

Managing Director
UPMSCL