

Technical Evaluation report against tender ref.no.-UPMSCL/EQ/RC/334 for Centrifuge Machine.

Sr. no.	Name of the Schedule	Conclusion of the received bids
1.	Centrifuge Machine	Qualified bids: <ol style="list-style-type: none">1. M/s. Labtop2. M/s. Remi Elektrotechnik Limited3. M/s. Esteem Industries Inc. Disqualified: <ol style="list-style-type: none">1. M/s. Biotechnologies Inc. Reasons: <ol style="list-style-type: none">a. Not Submitted Quality certificates like: ISI/CE marked/USFDA and ISO 13485 certificates even after clarification asked.b. Not submitted Performance certificate for same quoted model (atleast 10% of the schedule quoted).c. Not Submitted sufficient Purchase order issued to M/s. Biotechnologies Inc. (atleast 50% of the tender quantity i.e. 850 units) for supplying similar equipments.

N.B.: If there is any representation against above mentioned technical evaluation report kindly submit it by 20.02.2021 by 2:00 P.M., any representation submitted after above mentioned time period will not be entertained.