

Technical Evaluation report against tender ref. no.-UPMSCL/EQ/RC/472 for Oxygen generation Plant.

Sr. no.	Name of the Schedule	Conclusion of the received bids
1.	Oxygen Generation Plant	<p>Qualified bids:</p> <ol style="list-style-type: none">1. M/S. CBS Technologies Pvt. Ltd2. M/S. Landsky Engineers Pvt. Ltd3. M/S. Neometrix Engineering Pvt. Ltd <p>Disqualified:</p> <ol style="list-style-type: none">1. M/s. All Time Data Pvt. Ltd<p>Reason:</p><ol style="list-style-type: none">a. Not Submitted Quality Certificate.(US FDA approved or CE or should have CE(Conformity European)/European Certificate(EC)2. M/S. JP Singhal & Company<ol style="list-style-type: none">a. Not Submitted purchase order for executing similar work with satisfactory performance certificate for same order.b. Did not submit ISO 10083/ EN ISO 7396-1/ EN 737-3 European Standards and should be in Accordance with medical device directives 93/42/EEC.3. M/S. Kashi Pharmaceutical<ol style="list-style-type: none">a. Did not submit the OEM Authorization certificate.b. Not submitted purchase order for executing similar work with satisfactory performance certificate for same order.c. Not Submitted Quality Certificate. .(US FDA approved or CE or should have CE(Conformity European)/European Certificate(EC)4. M/S. MAG5 Innovations Pvt. Ltd<ol style="list-style-type: none">a. Not Submitted Annual Turnover for last three financial years.b. Not Submitted purchase order for executing similar work with Satisfactory performance certificate for same order.c. Did Not Submit ISO 10083/ EN ISO 7396-1/ EN737-3 European Standards and should be in Accordance with medical device directives

		<p>93/42/EEC.</p> <p>5. M/S. Medicare Plus Healthcare Pvt. Ltd.</p> <ul style="list-style-type: none">a. Not submitted Annual turnover for last 3 financial years.b. Not Submitted Single purchase order for executing similar work with Satisfactory performance certificate for same orderc. Not Submitted ISO 10083/ EN ISO 7396-1/ EN 737-3 European Standards and should be in Accordance with medical device directives 93/42/EEC.d. Not submitted US FDA approved or CE or should have CE(Conformity European)/European Certificate(EC) <p>6. M/S. Paras Defence and Technologies Ltd</p> <ul style="list-style-type: none">a. Not Submitted Annual Turnover for last three financial years.b. Not Submitted Single purchase order for executing similar work with satisfactory performance certificate for same order.c. Not Submitted ISO 13485: 2016 Certificationd. Not submitted ISO 10083/ EN ISO 7396-1/ EN 737-3 European Standards and should be in Accordance with medical device directives 93/42/EEC. <p>7. M/S. Suntech Industries</p> <ul style="list-style-type: none">a. Not Submitted ISO certification or GMP Certificateb. Not submitted ISO 10083/ EN ISO 7396-1/ EN 737-3 European Standards and should be in Accordance with medical device directives 93/42/EEC.c. Not submitted US FDA approved or CE or should have CE(Conformity European)/European Certificate(EC) <p>8. M/S. ATR Enterprises</p> <ul style="list-style-type: none">a. Did not Not Submit EMD (Submission of EMD is required as per Tender conditions). <p>9. M/S. Care Systems</p> <ul style="list-style-type: none">a. Did not Submit EMD(Submission of EMD is required as per tender conditions)
--	--	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Dated: 4th August , 2021

		<p>10. M/S. Lifesaving Inc. a. Did not Submit EMD(Submission of EMD is required as per tender conditions).</p> <p>11. M/S Colt Equipments Pvt. Ltd. a. Did not submit EMD(Submission of EMD is required as per Tender Conditions)</p>
--	--	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

N.B.: If there is any representation against above mentioned technical evaluation report kindly submit it by 5.08.2021 by 5:00 P.M., any representation submitted after the mentioned time period will not be entertained.