



## UTTAR PRADESH MEDICAL SUPPLIES CORPORATION LTD.

(A Government of Uttar Pradesh Undertaking)

Regd. Office: SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226 010

### Corrigendum-3 dated 24.12.2022

With reference to tender no. **UPMSCL/Drugs-179/019** dated 09 November, 2022 a corrigendum is being issued as follows:

S.N.	Annexure-A Schedule of requirement	Existing	Revised/ clarification
1	RDT (HEPA-C) Rapid Diagnostic Kit For Hepatitis-c	Kit needs to be able to detect both antibodies (capsid, NS3 and NS4) and antigen (Capsid) against/of Hepatitis C Virus.	Kit needs to be able to detect <b>antibodies</b> (capsid, NS3 and NS4) against/ of Hepatitis C virus.
2		The assay should have sensitivity of > 99 % and specificity of > 99%.	The assay should have sensitivity of more than or equal to 99% and <b>specificity of more than or equal to 98%</b> .
3	RDT (HEPA-C) Rapid Diagnostic Kit For Hepatitis-c	Required Quantity (in no. of Unit/test)- 4,00,000	Required Quantity (in no. of Unit/test)- <b>21,24,000</b>
4	RDT (HEPA-B) Rapid Diagnostic Kit For Hepatitis-B	Required Quantity (in no. of Unit/test)- 4,00,000	Required Quantity (in no. of Unit/test)- <b>21,24,000</b>

All other specifications of items and terms & conditions of the tender document shall remain same.

**MANAGING DIRECTOR**