



Uttar Pradesh Medical Supplies Corporation Limited

(A Govt. of Uttar Pradesh Undertaking)

GSTIN: 09AACCU2250P1ZZ CIN: U85310UP2018SGC102425

Registered Office : SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226002

E-Mail: drugs@upmsc.in Website: www.upmsc.in Contact Number: 0522-2838102

Corrigendum-7 Dated 18/11/2024

With reference to tender no. **UPMSCL/Drugs-217/09** dated 28.08.2024, a corrigendum is being issued as follows:

Date Corrigendum with Technical Specification of Rapid Diagnostic Kit for Hepatitis -B.

2) HBsAg (Rapid Test)

1. Should be coated with monoclonal antibodies covering all subtypes and variants of HBsAg
2. The assay should be able to detect surface antigen to Hepatitis B virus.
3. Should be compatible with plasma and serum both.
4. Adequate documents detailing the principle, components, biosafety, methodologies, validity criteria, interpretation of results, performance characteristics, storage conditions, limitation of assays, manufacturing and expiry dates should be provided with each kit.
5. The kit should have approval of the statutory authority from the country of origin
6. In case of imported kits it should be registered and licensed by the DCG(I)
7. In case of indigenous manufactures should be licensed by the competent authority/Licensing authority defined under Drugs and Cosmetics Act (1940) and medical device rule 2017
8. The kit should have minimum shelf life of 60% or 12 months(whichever is more) at the port/place of discharge of consignees.
9. The total procedure time shall not be more than 30 minutes.
10. The assay components should include positive and negative controls sufficient for conducting 20% of the tests (10% negative and 10% positive controls) which may be provided along with the kits if not a part of the kit.
11. The assay should have sensitivity of 100%and specificity of more than or equal to 98% as per the office order of MoHFW vide F. No 29/Misc./4/2016-DC(65) dated 13/6/2017
12. The control dot/band should be able to detect the presence of human immunoglobulin and should not be just a "procedural control" or meant merely for checking the flow of reagents or integrity of antigens except lateral flow technology

General Specifications

1. The manufacturer /authorized agent should ensure maintenance of cold chain during storage and transport of kits at 2-8^o C. The cumulative time temperature indicator technology should be used on each kit and be pre-qualified by WHO.
2. The pack size should not be more than 50 tests wherein each test is individually packed.
3. 8 kits should be supplied along with the procurement lot of which four kits will be used for validation, subject to which the kits of each batch and lot no. will be supplied to program and four kits will be retained for evaluation close to the expiry of the kit or in case of legal dispute with regards to performance of the kit as per specified parameters
4. The kit will be evaluated on the above parameters by the centers approved by the program

The committee approved the specification of HBsAg (rapid test)

S.N	Reference of tender document	Existing date	Revised date
1	Last Date and Time for Online Submission of Tender	19 th November,2024 UPTO 15:00 Hrs.	25 th November,2024 UPTO 15:00 Hrs.
2	Date and Time of Opening of Technical BID-COVER 'A'	19 th November,2024 UPTO 15:30 Hrs.	25 th November,2024 UPTO 15:30 Hrs.

All other terms & conditions of the tender document shall remain same.

**MANAGING DIRECTOR
UPMSCL**