



Uttar Pradesh Medical Supplies Corporation Limited

(A Govt. of Uttar Pradesh Undertaking)

(CIN: U85310UP2018SGC102425)

Registered Office : SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226010

E-Mail: equipment@upmsc.in Website: www.upmsc.in Contact Number: 0522-2838102

Date- 16.11.2022

Tender Reference No.:- GEM/2022/B/2578993

Technical Corrigendum

Name of Equipment: ICU Bed

Sn. NO.	Reference Of Tender Clause	Tender Clause	Revised Clause
1	Technical Specification Point 1	Exit Area: both sides of the bed should be illuminated to help & improve resident / patient safety security.	To be deleted
2	Technical Specification Point 3	Standard Length of approx: 2200-2350 mm (215-220 cm).	Standard Length of approx: 2200-2350 mm (215-220 cm)±10mm
3	Technical Specification Point 7	It should have very smooth mobility with the High Grade Castors mainly: Imported, with having Standard Diameter between: 120-130 mm & with Locking facility: diagonal for safe shifting of patient for performing routine fumigation of the Intensive Care Unit, along with Central Locking System/CPR System & Nurse Control System.	It should have very smooth mobility with the High Grade Castors mainly: having Standard Diameter between: 120-130 mm & with Locking facility for safe shifting of patient for performing routine fumigation of the Intensive Care Unit, along with Central Locking System/CPR System & Nurse Control System. Mechanism . Castor wheel mechanism required is Central locking system/Diagonal wheel mechanism
4	Technical Specification Point 8	Standard and Safety: a. CE(EU) from a notified Body/BIS b. ISO 13485:2016 c. IEC-60601-1	Standard and Safety: a. CE certificate for product from any European Agency/USFDA registration certificate
	Delivery Days	Delivery Days-30	Delivery Days-60 Days
5	Additional Point		Side rail should be drop down and detachable & made of ABS plastic
6	Additional Point		Mattress should be of minimum 4 inch thick ,foam of minimum 40 density with superior cover material(fluid resistant, antibacterial & flame retardant)
	Additional Point		The bidder has to quote according to the specifications attached in buyer uploaded ATC
	Additional Point		Delivery location will be multiple i.e. various Health Facilities of Govt. of UP.
	Additional Point		Not more than 5 days will be given to prepare sample &



Uttar Pradesh Medical Supplies Corporation Limited

(A Govt. of Uttar Pradesh Undertaking)

(CIN: U85310UP2018SGC102425)

Registered Office : SUDA Bhawan, 7/23, Sector-7, Gomti Nagar Extension, Lucknow-226010

E-Mail: equipment@upmsc.in Website: www.upmsc.in Contact Number: 0522-2838102

			transportation for sample demonstration after technical evaluation
	Additional Point		Payment terms will be as per UPMSCL payment terms & condition
	Additional Point		Past Performance will be as per tender terms & condition